

Certificate of Compliance

Certificate Number 20101104-E215059
Report Reference E215059-A2-UL, 2010 September 17
Issue Date 2010 November 4

Page 1 of 1



Issued to: **EPPENDORF A G**

BARKHAUSENWEG 1
22339 HAMBURG GERMANY

*This is to certify that
representative samples of*

Laboratory Use Electrical Equipment
Centrifuge
Vacufuge plus and Vacufuge plus System, model 5305

Have been investigated by Underwriters Laboratories Inc.® (UL) or any authorized licensee of UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1, Second Edition
CAN/CSA-C22.2 No. 61010-1, Second Edition
IEC 61010-2-010 - Edition 2 - Issue Date 2003/06/01
IEC 61010-2-020 - Edition 2 - Issue Date 2006/05/01

Additional Information: See UL On-Line Certification Directory at www.UL.com for additional information

Only those products bearing the UL Listing Mark for the US and Canada should be considered as being covered by UL's Listing and Follow-Up Service meeting the appropriate requirements for US and Canada.

The UL Listing Mark for the US and Canada generally includes: the UL in a circle symbol with "C" and "US" identifiers:

 the word "LISTED"; a control number (may be alphanumeric) assigned by UL; and the product category name (product identifier) as indicated in the appropriate UL Directory.

Look for the UL Listing Mark on the product

William R. Carney Director,

North American Certification Programs

Underwriters Laboratories Inc.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.